

Electrical and pneumatic analogue chart recorders for use in industrial process control systems —

Part 2: Guidance for inspection and routine testing

The European Standard EN 60873-2:2004 has the status of a
British Standard

ICS 25.040.40

National foreword

This British Standard is the official English language version of EN 60873-2:2004. It is identical with IEC 60873-2:2004.

The UK participation in its preparation was entrusted by Technical Committee GEL/65, Measurement and control, to Subcommittee GEL/65/2, Elements of systems, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the *BSI Catalogue* under the section entitled “International Standards Correspondence Index”, or by using the “Search” facility of the *BSI Electronic Catalogue* or of British Standards Online.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 12 May 2004

Summary of pages

This document comprises a front cover, an inside front cover, the EN title page, pages 2 to 6, an inside back cover and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

Amendments issued since publication

Amd. No.	Date	Comments

© BSI 12 May 2004

ISBN 0 580 43772 8

EUROPEAN STANDARD

EN 60873-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2004

ICS 25.040.40

English version

**Electrical and pneumatic analogue chart recorders
for use in industrial process control systems
Part 2: Guidance for inspection and routine testing
(IEC 60873-2:2004)**

Enregistreurs de courbes électriques
et pneumatiques pour une utilisation
dans les systèmes de processus
industriels
Partie 2 : Guide pour les inspections
et les essais individuels
(CEI 60873-2:2004)

Elektrische und pneumatische analoge
Streifenschreiber zum Einsatz in
Systemen industrieller Prozessleittechnik
Teil 2: Anleitung für die Abnahme
und Betriebsprüfung
(IEC 60873-2:2004)

This European Standard was approved by CENELEC on 2004-03-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

© 2004 CENELEC - All rights of exploitation in any form and by any means reserved worldwide for CENELEC members.

Ref. No. EN 60873-2:2004 E

Foreword

The text of document 65B/513/FDIS, future edition 1 of IEC 60873-2, prepared by SC 65B, Devices, of IEC TC 65, Industrial-process measurement and control, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60873-2 on 2004-03-01.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 2004-12-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 2007-03-01

Annex ZA has been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60873-2:2004 was approved by CENELEC as a European Standard without any modification.

ELECTRICAL AND PNEUMATIC ANALOGUE CHART RECORDERS FOR USE IN INDUSTRIAL PROCESS CONTROL SYSTEMS –

Part 2: Guidance for inspection and routine testing

1 Scope and object

This part of IEC 60873 applies to electrical and pneumatic analogue chart recorders (for use in industrial-process control systems), operating from a standardized signal which may be used in process control. It is intended that continuous and dotted line traces, and multiple pen and multiple-channel instruments should be covered. Some tests may not apply to all instruments and additional tests may be required for certain types of recorders.

The object of this standard is to provide technical guidance for inspection and routine testing of electrical and pneumatic analogue chart recorders, for instance, as acceptance tests or after repair. For a full evaluation, or where a recorder is to be used under arduous conditions, (for example, high temperatures), IEC 60873-1 should be used. Whenever possible any tests carried out are to be in accordance with IEC 61298, and the person carrying out the tests needs to have a copy of it to hand and be familiar with it.

Quantitative criteria for acceptable performance should be established by agreement between manufacturer and user, and the report on the tests shall make clear which tests were carried out. The requirements of this standard shall be effective when agreed by the manufacturer and the user.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60410:1973, *Sampling plans and procedures for inspection by attributes*

IEC 60873-1, *Electrical and pneumatic analogue chart recorders for use in industrial-process control systems – Part 1: Methods for performance evaluation*¹

IEC 61010-1:2001, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements*

IEC 61298 (all parts), *Process measurement and control devices – General methods and procedures for evaluating performance*

¹ To be published.

3 Terms and definitions

For the purposes of this document, the terms and definitions of IEC 60873-1, together with the following, apply.

3.1

acceptance test

test which intends to prove to the user that the device complies with certain conditions of its requirements as they appear in the contract

4 Sampling for test

If, by agreement between user and manufacturer, tests are to be performed on a sample lot, it is recommended that a sampling method such as that presented in IEC 60410 be selected.

When sampling is used, recorders to be tested may be chosen by the user's inspector.

5 Performance test

Environmental conditions at the location of testing shall be recorded (see 5.1 of IEC 60873-1).

Adjustments: For zero, span and damping, the routine tests shall be carried out with the final manufacturing adjustments (for example, as an acceptance test) or with the adjustments determined by the user (for example, after repair).

The following tests shall be performed unless another series of tests is agreed between user and manufacturer.

NOTE Ink and paper should be in accordance with the manufacturer's specifications. A chart speed of approximately 20 mm/h should preferably be used.

5.1 Measured error and hysteresis (see Clause 6 of IEC 60873-1)

Prior to recording observations, the device under test shall be exercised by three full-range traverses in each direction. The input-output characteristic under reference conditions shall be measured in one measurement cycle traversing full range in each direction. For this, at least five points of measurements should be distributed over the range, for example, in steps of 25 % from 0 % to 100 % and then from 100 % to 0 %. Tapping or vibrating the instrument under test is not allowed, unless otherwise stated.

Adjust the input to bring the output to the required value. Measure the input at this level and determine the difference between this input and the specified input to give this output value. As the result, report the difference as "measured error including hysteresis" expressed in per cent of input span.

NOTE For this test, measured error is defined as the greatest positive or negative error in the measuring cycle. Hysteresis (including deadband) is defined as the greatest difference between upscale and downscale readings which are determined at each value of output.

5.2 Effects of influence quantities

5.2.1 Power supply variations (see 12.1 of IEC 61298-3)

Adjust the input signal to bring the output to approximately 95 % of span. Measure and report the change of input that is necessary to bring the output level to its original value in % of span at the following variations in power supply or at the manufacturer's stated limits, if smaller:

- voltage variation = +10 %, –15 % of nominal a.c. or +20 %, –15 % of nominal d.c. voltage;
- supply pressure variation = +10 % –15 % of nominal supply pressure (see 12.8 of IEC 61298-3)

5.2.2 Over-range (see Clause 10 of IEC 61298-3)

With the input adjusted to bring the output first to approximately 5 % and then to approximately 95 % span, note the input levels.

The input shall then be increased gradually to the maximum overrange specified by the manufacturer. After the over-range has been applied for 1 min, the input shall be reduced to the nominal lower range value. After a further 5 min have elapsed, using the same output values as before, determine the changes required in input values to bring the output back to the originally used values.

The changes in input determined after over-ranging shall be reported, expressed as per cent input span.

NOTE If over-ranging produces significant thermal effects, the duration of application should be increased accordingly.

5.3 Recording quality

The instrument shall be connected as for normal operation and an alternating input applied with a peak-to-peak amplitude equal to half the span and centred at the mean of the upper and lower range input values. The frequency selected shall be such that all recorded traces can be clearly distinguished (not more than 1 cycle per millimeter chart travel).

After an appropriate number of traverses it shall be noted

- a) whether all traces are without interruption of the ink flow;
- b) whether the widths of the recorded lines change during the test;
- c) for multi-channel recorders, whether the ink colours change over a longer distance than 5 mm after crossing different colour traces.

5.4 Paper speed

With the chart-drive operating as normal, check by observing over a suitable period of time that the paper speed is within specification.

5.5 Step response

With the chart being driven at its fastest selectable speed, apply a suitably large step change of input signal and check that the step response of the pen mechanism is within specification.

6 Safety

The instrument shall be checked against the requirements of IEC 61010-1.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE Where an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60410	1973	Sampling plans and procedures for inspection by attributes	-	-
IEC 60873-1	- 1)	Electrical and pneumatic analogue chart recorders for use in industrial-process control systems Part 1: Methods for performance evaluation	EN 60873-1	2004 2)
IEC 61010-1	2001	Safety requirements for electrical equipment for measurement, control, and laboratory use Part 1: General requirements	EN 61010-1 + corr. June	2001 2002
IEC 61298	Series	Process measurement and control devices - General methods and procedures for evaluating performance	EN 61298	Series

1) Undated reference.

2) Valid edition at date of issue.

.....

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001. Fax: +44 (0)20 8996 7001. Email: orders@bsi-global.com. Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: +44 (0)20 8996 7111. Fax: +44 (0)20 8996 7048. Email: info@bsi-global.com.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: +44 (0)20 8996 7002. Fax: +44 (0)20 8996 7001. Email: membership@bsi-global.com.

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsi-global.com/bsonline>.

Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright & Licensing Manager. Tel: +44 (0)20 8996 7070. Fax: +44 (0)20 8996 7553. Email: copyright@bsi-global.com.